




Salinity Sensor Adapter Datasheet



1 Introduction

The LDSBus Salinity Sensor Adapter is designed to work with a Salinity probe to form a complete Salinity sensor. It uses electrical conductivity method for its measurement. This adapter has a BNC connector for attaching the Salinity probe. A 2-point calibration method is used to calibrate the adapter and probe and allows measurements of salinity between 1ppt and 120ppt with a precision of 1ppt. This adapter and probe is suitable for use in applications such as aquaculture fishponds, shrimp ponds, sea water applications, nutrient tanks, water treatment plants, and swimming pools.

1.1 Features

- Salinity Sensor Adapter to integrate with cell constant $K=10$ Salinity probe with BNC connector
- Measures Salinity range of 1ppt to 120ppt with linearized output and 1ppt resolution
- 2 Point calibration
- Supports BRTSys LDSBus protocol
- High report rate of every 5 seconds
- Low power consumption: 91mW at 5V
- Operating temperature range: 0°C to +70°C
- Flush Mount and DIN Rail Mount options
- Supported platforms:
 - IoTPortal
 - LDSBus Python SDK
 - LDSBus .Net SDK

Visit <https://brtsys.com/resources> for more information.



Neither the whole nor any part of the information contained in, or the product described in this manual, may be adapted, or Reproduced in any material or electronic form without the prior written consent of the copyright holder. This product and its documentation are supplied on an as-is basis and no warranty as to their suitability for any particular purpose is either made or implied. BRT Systems Pte Ltd (BRTSys) will not accept any claim for damages howsoever arising as a result of use or failure of this product. Your statutory rights are not affected. This product or any variant of it is not intended for use in any medical appliance, device, or System in which the failure of the product might reasonably be expected to result in personal injury. This document provides preliminary information that may be subject to change without notice. No freedom to use patents or other intellectual property rights is implied by the publication of this document. BRT Systems Pte Ltd, 1 Tai Seng Avenue, Tower A, #03-01 Singapore 536464. Singapore Registered Company Number: 202220043R.

2 Part Numbers

Part#	Naming
LS-1301-01A	LDSBus Salinity Sensor Adapter
LA-0501-01A	LDSBus RJ11-RJ11 Cable (5m)
LA-1201-01A	LDSBus DIN Rail Mount Set

Table 1 - Part Numbers / Ordering Information

Table of Contents

1	Introduction	1
2	Part Numbers	2
3	Specifications	4
4	FCC Compliance Statement.....	5
5	Hardware Features.....	6
6	Sensor Adapter Configuration and Installation	8
6.1	Connection Diagram.....	8
7	Mounting Instructions.....	9
7.1	Flush Mount	9
7.2	DIN Rail Mount.....	9
8	Mechanical Dimensions	10
9	System Status LED Indicators	11
10	Probe Selection	12
11	Contact Information.....	13
Appendix A	– References	14
Document References	14
Acronyms and Abbreviations	14
Appendix B	– List of Figures and Tables.....	15
List of Figures	15
List of Tables	15
Appendix C	– Revision History	16

3 Specifications

Features	Interface	BNC (Connect to Salinity probe) RS485
	LED Indicator (RGB)	System Status Indicator (Please refer to LED section)
	Mounting	Flush Mount
DIN Rail Mount		
Power	Device Input Voltage	5V DC Bus Power
	Typical Power	5V 91mW
	Max. Power	266mW
Salinity Sensor input module	Detection Range	1 – 120ppt
	Resolution	0.01ppt
	Response Time	<1Minute
	Calibration	2 Point Calibration
Physical Characteristics	Color	White
	Housing	Polycarbonate
	Dimensions	L117.6mm x W42.9mm x H29.7mm
Environmental Limits	Operating Temperature	0 to 70°C
	Storage Temperature	-20 to 85°C
	Ambient Relative Humidity	5 to 95% (non-condensing)
Package Contents	Device	1x LDSBus Salinity Sensor Adapter
	Wire Assembly	1x LDSBus RJ11-RJ11 Cable (5m)
Optional	Mounting Accessories	1x LDSBus DIN Rail Mount set

Table 2 - LDSBus Salinity Sensor Adapter Specifications

4 FCC Compliance Statement

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions:

- (1) These devices may not cause harmful interference, and
- (2) These devices must accept any interference received, including interference that may cause undesired operation.

NOTE: The equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If the equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Re-orient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

To maintain compliance with FCC's RF exposure guidelines, at least 20cm of separation distance between the device and the user's body must be always maintained.

FCC Radiation Exposure Statement

This device complies with FCC radiation exposure limits set forth for an uncontrolled environment and it also complies with Part 15 of the FCC RF Rules. This equipment must be installed and operated in accordance with provided instructions and the antenna(s) used for this transmitter must be installed to provide a separation distance of at least 20 cm from all persons and must not be co-located or operating in conjunction with any other antenna or transmitter. End-users and installers must be provided with antenna installation instructions and consider removing the no-collocation statement.

Caution

Any changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.



5 Hardware Features




Figure 1 - LDSBus Salinity Sensor Adapter Hardware Features

Function	Labels	Description
BNC Connector	Probe	Probe Interface
Salinity Electrode Probe	-	Sensing Probe
System Status LED Indicator	LED	LDSBus status LED
RJ12 Port	LDSU	LDSBus data and power interface port. The physical port is RJ12. The connection interface can be RJ11/RJ12.

Table 3 - LDSBus Salinity Sensor Adapter Hardware Features

6 Sensor Adapter Configuration and Installation

Please refer to [LDSBus Configuration Utility User Guide](#) on how to configure the device name, address, and termination settings before using it for your application.

6.1 Connection Diagram

Figure 2 illustrates the connection of the LDSBus Salinity Sensor Adapter (LDSBus Device) to the LDSBus. Please visit <https://brtsys.com/resources> to view the full device application, setup, and installation guides.




Figure 2 - LDSBus Salinity Sensor Adapter to LDSBus - Connection Diagram

Setup Instructions:

1. Connect the first LDSBus Quad T-Junction to any of the LDSBus Host Systems using the RJ45 (CAT5e) cable as shown in Figure 2.
2. Connect the configured LDSBus Salinity Sensor Adapter to the LDSBus Quad T-Junction as shown in Figure 2.
3. If there are more than one LDSBus Quad T-Junctions, chain them together as shown in Figure 2.

Enable termination for the last device in LDSBus.

7 Mounting Instructions

7.1 Flush Mount

The device can be flush mounted directly on a wall or any flat surface using 2 M3.5*16mm (thread) screws.




Figure 3 - LDSBus Salinity Sensor Adapter Flush Mount

7.2 DIN Rail Mount

The device can be mounted on a DIN Rail using the LDSBus DIN Rail Mount set. This set is optional and includes the bracket and mounting screws. Refer to Table 1 for more details.




Figure 4 - LDSBus Salinity Sensor Adapter DIN Rail Mount

8 Mechanical Dimensions




Figure 5 - LDSBus Salinity Sensor Adapter Dimension – Top View




Figure 6 - LDSBus Salinity Sensor Adapter Dimension – Side View




Figure 7 - LDSBus Salinity Sensor Adapter Dimension – Bottom View

Note: All dimensions are in millimetres.

9 System Status LED Indicators

LDSU devices come with a tri-color LED, and LED status colors are mentioned in the table below.

Status display colors:

- | | | | |
|----|--------|---|--|
| 1. | RED | - | Device in error conditions |
| 2. | YELLOW | - | Un-configured device |
| 3. | GREEN | - | Device in normal state (Device termination is OFF) |
| 4. | BLUE | - | Device in normal state (Device termination is ON) |

Device Status	LED Color		Flashing Frequency	Description
Un-configured device	YELLOW		LED flashing @1Hz	Un-configured device with factory default address (126)
Configured device	GREEN		Steady-Non-flashing	Configured device (Device ID 1-125) and device is idle.
	BLUE			
Addressed device	GREEN		LED flashing @5Hz	Device is busy communicating.
	BLUE			
Identified device	GREEN		LED flashing @1Hz	Device in identify state.
	BLUE			
Device error	RED		Steady - Non-flashing	Device error has occurred.
Firmware update	YELLOW		Steady - Non-flashing	Device firmware update.

Table 4 - System Status LED Indicators

10 Probe Selection

The following specifications are recommended for selecting a Probe -

Detection Range	Cell Constant	Connector
1ppt to 120ppt	K=10	BNC

Table 5 - Probe Specifications

For more information on calibration, please refer to the section "Calibration Procedure" in [LDSBus Configuration Utility User Guide](#).

For information related to probes recommendation and selection criteria, please refer to [3rd Party Compatible Probes Specifications](#).

11 Contact Information

Refer to <https://brtsys.com/contact-us/> for contact information.

System and equipment manufacturers and designers are responsible to ensure that their systems, and any BRT Systems Pte Ltd (BRTSys) devices incorporated in their systems, meet all applicable safety, regulatory and system-level performance requirements. All application-related information in this document (including application descriptions, suggested BRTSys devices and other materials) is provided for reference only. While BRTSys has taken care to assure it is accurate, this information is subject to customer confirmation, and BRTSys disclaims all liability for system designs and for any applications assistance provided by BRTSys. Use of BRTSys devices in life support and/or safety applications is entirely at the user's risk, and the user agrees to defend, indemnify, and hold harmless BRTSys from any and all damages, claims, suits, or expense resulting from such use. This document is subject to change without notice. No freedom to use patents or other intellectual property rights is implied by the publication of this document. Neither the whole nor any part of the information contained in, or the product described in this document, may be adapted, or reproduced in any material or electronic form without the prior written consent of the copyright holder. BRT Systems Pte Ltd, 1 Tai Seng Avenue, Tower A, #03-01, Singapore 536464. Singapore Registered Company Number: 202220043R.

Appendix A – References

Document References

[LDSBus Configuration Utility User Guide](#)

[LDSBus Python SDK V3.0.0 Guide](#)

[LDSBus .Net SDK V3.0.0 Guide](#)

[3rd Party Compatible Probes Specifications](#)

[Sensors and Actuators Quick Start Guide for USB Hosts](#)

[Sensors and Actuators Quick Start Guide for IoTPortal](#)

Acronyms and Abbreviations

Terms	Description
DC	Direct Current
LDSBus	Long Distance Bus Sensor
LED	Light Emitting Diode

Appendix B – List of Figures and Tables

List of Figures

Figure 1 - LDSBus Salinity Sensor Adapter Hardware Features	6
Figure 2 - LDSBus Salinity Sensor Adapter to LDSBus - Connection Diagram	8
Figure 3 - LDSBus Salinity Sensor Adapter Flush Mount	9
Figure 4 - LDSBus Salinity Sensor Adapter DIN Rail Mount	9
Figure 5 - LDSBus Salinity Sensor Adapter Dimension – Top View	10
Figure 6 - LDSBus Salinity Sensor Adapter Dimension – Side View	10
Figure 7 - LDSBus Salinity Sensor Adapter Dimension – Bottom View	10

List of Tables

Table 1 - Part Numbers / Ordering Information	2
Table 2 - LDSBus Salinity Sensor Adapter Specifications	4
Table 3 - LDSBus Salinity Sensor Adapter Hardware Features	7
Table 4 - System Status LED Indicators	11
Table 5 - Probe Specifications	12

Appendix C – Revision History

Document Title: LDSBus Salinity Sensor Adapter Datasheet
 Document Reference No.: BRTSYS_000032
 Clearance No.: BRTSYS#028
 Product Page: <https://brtsys.com/product/salinity-sensor-adapter/>
 Document Feedback: [Send Feedback](#)

Revision	Changes	Date
Version 1.0	Initial Release	26-01-2023
Version 1.1	Updated the following: HVT references to Quad T-Junction; Singapore address	11-09-2023
Version 1.2	Section 1.1 - Added .Net SDK to Supported platforms; ROHS icon added Section 2 - Added LA-0501-01A in part number Section 3 - Updated "Wire Assembly" Section 4 - Added FCC statement Section 5 - Added Table 3; Updated Figure 1 (Changed from "RJ11" to RJ12" for LDSU port) Section 6.1 - Added point #3 & #4 Appendix A – References > Document References - updated	23-01-2026